

GREEN&CLEAN WK

Application using the dosing aid



1. Wear gloves when removing the dosing aid from its packaging



2. Remove the warranty strip (avoid touching the pick-up tube)



3. Insert pick-up tube and screw dosing aid tightly on the GREEN&CLEAN WK bottle*



4. Open the yellow cap and press lightly to dose (see dosing table below)



5. Pour GREEN&CLEAN WK into the water bottle and reconnect it to the dental unit



6. Close cap immediately after every use to protect contents from contamination

2% hydrogen peroxide solution for the decontamination of process water and the prevention of limescale deposits in the pipe systems of dental units

Characteristics broad spectrum of efficacy, with long-term effect
prevention of biofilm
limescale inhibitor prevents calcification
protects hand and angle pieces
CE 0044

Recommendations for use wear protective gloves and glasses
for continuous decontamination, add GREEN&CLEAN WK every time the water bottle is refilled (see dosing table below)
when the contents of the GREEN&CLEAN WK bottle have been used up, remove the dosing aid from the empty bottle and transfer it directly to the new bottle
replace the dosing aid every six months and in case of heavy contamination the dosing aid is not suitable for reprocessing
*the dosing aid is only compatible with the 750 ml GREEN&CLEAN WK bottle - use a measuring beaker when using 1 liter GREEN&CLEAN WK bottles



Dosing table

Water	GREEN&CLEAN WK	
	Standard decontamination	Shock treatment**
0.5 l	6 ml	12 ml
1.0 l	12 ml	24 ml
2.0 l	24 ml	48 ml
3.0 l	36 ml	72 ml

**action time for shock treatment: min. 1 hour, max. 12 hours (tubes must be soaked completely for the entire duration of the shock treatment)